

ROY COOPER
Governor

MICHAEL S. REGAN
Secretary

LINDA CULPEPPER
Director



NORTH CAROLINA
Environmental Quality

August 14, 2019

Christel Compton
Program Manager
The Chemours Company FC, LLC
Fayetteville Works
22828 NC Highway 87 W
Fayetteville, North Carolina 28306-7332

RE: Chemours Consent Order Paragraph 11a
Non-targeted Analysis and Characterization Sampling Plans

Dear Ms. Compton,

This is to confirm approval of the of the subject work plans for paragraph 11a with the following exceptions:

- Lab standards solvated in water is not an acceptable approach and not approved based on current research. Therefore, Chemours must use a solvent based system for these activities including providing lab standards for DEQ and EPA use.
- Direct injection will be considered an acceptable method for an analyte if the Method Reporting Level (MRL) is at a level comparable to private laboratories, which currently is 5 ppt or less. If a direct injection method is being developed will not reach that level of quantification, Chemours will need to implement a solid phase extraction method which will reach the available MRL comparable to private laboratories.

If you have any questions or would like to discuss further, please contact me at (919) 707-9014 or Allen Martin at (919) 733-3908 ext. 236

Sincerely,

A handwritten signature in blue ink, appearing to read "Linda Culpepper".

Linda Culpepper

cc: Allen Martin
Kevin Garon, Chemours
Brian Long, Chemours
Lam-Wing Leung, Chemours



North Carolina Department of Environmental Quality | Division of Water Resources
512 North Salisbury Street | 1611 Mail Service Center | Raleigh, North Carolina 27699-1611
919.707.9000